

# Schedule

Issue date: 07 September 2018  
Valid until: 22 March 2020



## NO: SMM 505

(Issue 5, 07 September 2018 replacement of SMM 505 dated 12 June 2018)

Page: 1 of 7

**LABORATORY LOCATION:**  
(PERMANENT LABORATORY)



**BUILDTEST LABORATORY SDN. BHD.**  
**NO. 12, JALAN PS 8/1**  
**TAMAN PRIMA SELAYANG**  
**68100 BATU CAVES**  
**SELANGOR**  
**MALAYSIA**

**FIELD(S) OF TESTING:**

**MECHANICAL & NON-DESTRUCTIVE TEST**

**FIELD(S) OF CALIBRATION:**

**MASS**

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2005 (ISO/IEC 17025:2005).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

### SCOPE OF TESTING: MECHANICAL

Materials/ Products Tested	Type of Test/ Properties Measured/ Range of Measurement	Standard Test Methods/ Equipment/Techniques
Hardened Concrete	Compressive Strength Test (Cubes, Cores & Cylinders)	BS EN 12390-3: 2009 (Test at ambient conditions)
	Water Absorption of Concrete Specimens	BS 1881-122: 2011 (Test at ambient conditions)
	Density of Hardened Concrete	BS EN 12390-7: 2009 (Volume by Water Displacement Method)
Aggregates	Flakiness Index of Coarse Aggregates	BS 812 : Part 105 : Sect 105.1 : 1989
	Elongation Index of Coarse Aggregates	BS 812 : Part 105 : Sect 105.2 : 1990
	Aggregate Crushing Value (ACV)	BS 812 : Part 110 : 1990 (Test at Dry Conditions)
	Aggregate Impact Value (AIV)	BS 812 : Part 112 : 1990 Clause 7.1 (Test at Dry Conditions)
	Ten Percent Fines Value (TFV)	BS 812 : Part 111 : 1990 (Test at Dry Conditions)

Scan this QR Code or visit [www.ism.gov.my/cab-df/rectories](http://www.ism.gov.my/cab-df/rectories) for the current scope of accreditation

**NO: SAMM 505**(Issue 5, 07 September 2018 replacement  
of SAMM 505 dated 12 June 2018)**SCOPE OF TESTING: MECHANICAL**

<b>Materials/ Products Tested</b>	<b>Type of Test/ Properties Measured/ Range of Measurement</b>	<b>Standard Test Methods/ Equipment/Techniques</b>
Aggregates	Los Angeles Abrasion (LA) of Small-Size Coarse Aggregates	ASTM C131/C131M-14
	Los Angeles Abrasion (LA) of Large-Size Coarse Aggregates	ASTM C535-16
	Particle Size Distribution By Sieving Method	BS EN 933-1 : 2012
	Determination of Materials Finer than 75 µm by Washing	ASTM C 117-17 (Procedure A - Washing with Plain Water)
	Particle Density and Water Absorption of Coarse Aggregates	BS 812 : Part 2 : 1995 Clause 5.3 - Wire Basket Method (Test at Ambient Conditions)
	Particle Density and Water Absorption of Fine Aggregates	BS 812 : Part 2 : 1995 Clause 5.5 - Glass Jar Method (Test at Ambient Conditions)
Organic Impurities in Fine Aggregates for Concrete	ASTM C40/C40M-16	
Soil	Moisture Content	BS 1377 : Part 2 : 1990 Clause 3.2 (Oven-drying Method)
	Dry Density / Moisture Content Relationship of Soils by 2.5kg Rammer Method	BS 1377 : Part 4 : 1990 Clause 3.3
	Dry Density / Moisture Content Relationship of Soils by 4.5kg Rammer Method	BS 1377 : Part 4 : 1990 Clause 3.5
	Dry Density / Moisture Content Relationship of Granular Soils by Vibrating Hammer Method	BS 1377: Part 4: 1990 Clause 3.7
	Liquid Limit	BS 1377: Part 2: 1990 Clause 4.3
	Plastic Limit and Plasticity Index	BS 1377: Part 2: 1990 Clause 5
	Laboratory California Bearing Ratio (CBR)	BS 1377 : Part 4 : 1990 Clause 7

**NO: SAMM 505**(Issue 5, 07 September 2018 replacement  
of SAMM 505 dated 12 June 2018)**SCOPE OF TESTING: MECHANICAL**

Materials/ Products Tested	Type of Test/ Properties Measured/ Range of Measurement	Standard Test Methods/ Equipment/Techniques
Steel Reinforcing Bar	Tensile Tests for determination of: <ul style="list-style-type: none"> <li>- Yield strength</li> <li>- Tensile strength</li> <li>- Mass per meter</li> <li>- Percentage elongation after fracture</li> <li>- Percentage total extension at maximum force</li> </ul> Force range: up to 1500kN  Bend Test   Rebend Test	MS ISO 15630-1 : 2012 Clause 5  ISO 6892-1 : 2016  MS 146:2014   MS ISO 15630 – 1: 2012 Clause 6 ASTM E 290 – 14 MS 146:2014 ISO 7438:2016   MS ISO 15630 – 1 : 2012 Clause 7 MS 146 : 2014 Clause 7.3.5 MS 145 : 2014 Clause 7.2.5
Steel Wire	Tensile Tests for determination of: <ul style="list-style-type: none"> <li>- Mass per metre</li> <li>- Yield strength (determined from 0.2% proof strength)</li> <li>- Tensile strength</li> <li>- Tensile/yield strength ratio</li> <li>- Percentage total elongation at maximum force</li> </ul> Bend Test   Rebend Test	MS ISO 15630-1 : 2012 Clause 5  ISO 6892-1 : 2016  MS 144:2014   MS ISO 15630 – 1: 2012 Clause 6 ASTM E 290 – 14 MS 146 : 2014 ISO 7438 : 2016   MS ISO 15630 – 1 : 2012 Clause 7 MS 146 : 2014 Clause 7.3.5 MS 145 : 2014 Clause 7.2.5

**NO: SAMM 505**(Issue 5, 07 September 2018 replacement  
of SAMM 505 dated 12 June 2018)**SCOPE OF TESTING: MECHANICAL**

<b>Materials/ Products Tested</b>	<b>Type of Test/ Properties Measured/ Range of Measurement</b>	<b>Standard Test Methods/ Equipment/Techniques</b>
Steel Fabric	Tensile Tests for determination of: - Yield strength (determined from 0.2% proof strength) - Tensile strength - Tensile/yield strength ratio - Percentage total elongation at maximum force  Bend Test  Rebend Test	EN ISO 15630-2 : 2010 Clause 5  ISO 6892-1 : 2016  MS 145:2014  MS ISO 15630 – 1 : 2012 Clause 6 ASTM E 290 – 14 MS 146 : 2014 ISO 7438 : 2016  MS ISO 15630 – 1 : 2012 Clause 7 MS 146 : 2014 Clause 7.3.5 MS 145 : 2014 Clause 7.2.5
Reinforcement Couplers for Mechanical Splices of Bars	Tensile Strength	ISO 6892 – 1: 2016
Masonry Units	Compressive Strength Test (Clay, Calcium Silicate and Aggregate Concrete Masonry Units)  Water Absorption	BS EN 772-1 : 2011 + A1 : 2015  BS EN 772 – 7: 1998
Bitumen	Sampling Compacted Asphalt Mixtures for Laboratory Testing  Thickness or Height of Compacted Bituminous Paving Mixture Specimens	ASTM D 5361 / D 5361 M-16  ASTM D 3549 / D 3549 M-17 (Method A)

Scan this QR Code or visit [www.ism.gov.my/cab-df/rectories](http://www.ism.gov.my/cab-df/rectories) for the current scope of accreditation**Signatory(ies):**

1. **Ip Kwok Khuen**
2. **Tang Wei Luen**
3. **Ip Kar Mun**

**NO: SAMM 505**(Issue 5, 07 September 2018 replacement  
of SAMM 505 dated 12 June 2018)

Page: 5 of 7

**SCOPE OF TESTING: MECHANICAL****SITE: CATEGORY I**

Materials/ Products Tested	Type of Test/ Properties Measured/ Range of Measurement	Standard Test Methods/ Equipment/Techniques
Soil	In-situ California Bearing Ratio (CBR)  In-situ Density Test by Small Pouring Cylinder Method  In-situ Density Test by Large Pouring Cylinder Method	BS 1377 : Part 9 : 1990 Clause 4.3  BS 1377 : Part 9 :1990 Clause 2.1 – Small Pouring Cylinder Method  BS 1377 : Part 9 :1990 Clause 2.2 – Large Pouring Cylinder Method

**Signatory(ies):**

1. Ip Kwok Khuen
2. Tang Wei Luen

# Schedule

Issue date: 07 September 2018  
Valid until: 22 March 2020



## NO: SMM 505

(Issue 5, 07 September 2018 replacement  
of SMM 505 dated 12 June 2018)

Page: 6 of 7

### SCOPE OF TESTING: NON-DESTRUCTIVE TEST

### SITE TESTING: CATEGORY I

Materials/ Products Tested	Type of Test/ Properties Measured/ Range of Measurement	Standard Test Method/ Equipment/Technique
Hardened Concrete	Surface Hardness Testing by Rebound Hammer in the range of 20 to 55 rebound number, R	BS EN 12504 – 2 : 2012

### Signatory(ies):

1. Ip Kwok Khuen
2. Tang Wei Luen

Scan this QR Code or visit [www.ism.gov.my/cab-direktories](http://www.ism.gov.my/cab-direktories) for the current scope of accreditation

# Schedule

Issue date: 07 September 2018  
Valid until: 22 March 2020



MS ISO/IEC 17025

## NO: SAMM 505

(Issue 5, 07 September 2018 replacement  
of SAMM 505 dated 12 June 2018)

Page: 7 of 7

\* The expanded uncertainties are based on an estimated confidence probability of not less than 95% and have a coverage factor of k=2 unless stated otherwise

**SCOPE OF CALIBRATION: MASS**

**SITE CALIBRATION: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
Balance	Up to 30 kg	0.1 g	ASTM E 898 – 88 (Reapproved 2013 – Calibrated by Using Standard Weights)

### Signatory(ies):

1. Ip Kwok Khuen
2. Ip Kar Mun

Scan this QR Code or visit [www.ism.gov.my/cab-direktories](http://www.ism.gov.my/cab-direktories) for the current scope of accreditation